

Date: April 2nd 2012

Certificate No:11-05982-01

The Laboratory of Iwate Pharmaceutical Association

3-17-37, Kamidou, Morioka-city, Iwate-Prefecture, Japan

Zip: 020-0125 TEL : 81-19-641-4401 FAX : 81-19-641-4792

EXAMINATION CERTIFICATE

Client	GINGA KOGEN BEER CO.LTD.,
Name of Goods	ground water
Location of sampling points	3-647-1,Sawauchi Kaisawa,Nishi-waga town,Waga-gun,Iwate
Sampled by	Kiyoto Kanno
Date of Receipt	March 30th 2012
Date of Examination	March 30th 2012

This is to certify that the samples by the client have been duly examined by our laboratory and the results are as follows;


Result of Examination

Item	Result	Quantitation limit
Iodine-131	Not Detected	0.84 Bq/kg
Cesium-134	Not Detected	0.84 Bq/kg
Cesium-137	Not Detected	0.72 Bq/kg

Method of Analysis : γ -ray Spectrometry (Germanium semiconductor Detector)

Measurement period : 2500 seconds.


Director : Yoshikazu Miyate


Date: