



No. 96-217366-190420-69012

RADIOACTIVITY MEASUREMENT REPORT

Applicant Address	3-647-1,SawauchiKaisawa,Nishi-waga town,Waga-gun,Iwate,Japan 029-5703
Name of Company	GINGAKOGEN BEER CO.LTD

Kotobiken Medical Laboratories Inc.
Food Analysis Center.
3-24 Oroshi Kikuta-machi, Koriyama-shi Fukushima.
963-0547,Japan.
Phone 024-963-1005
Fax 024-963-1033

The following submitted sample was tested in our laboratory and analytical results are as given below

Receptionist day	April.19.2019
Receipt number	69012
Name of Goods	Komugi no Beer
Date of Production	April.3.2019

The contents of measurement	
Detector	Germanium detector
Measuring time	2000 sec.
Measuring date	April.20.2019
Measuring place	Food Analysis Center Tukuba office : 445-1 Kamiyokoba Tsukuba-shi Ibaraki,305-0854

Items examined	Results	Limit of detection
^{137}Cs	Not detection	0.8 Bq/kg
^{134}Cs	Not detection	0.7 Bq/kg
^{131}I	Not detection	0.6 Bq/kg

Date of Issue April.20.2019

Etsuo Ishikawa
Center director
Kotobiken Medical Laboratories Inc.
Food Analysis Center.

